



KERALA AGRICULTURAL UNIVERSITY

CARDAMOM RESEARCH STATION, PAMPADUMPARA

IDUKKI DISTRICT, KERALA, 685 553, Ph. 04868 236263

E-mail: crspam@kau.in

CRSP(2)-333/18

Dated:29.09.2018

QUOTATION NOTICE

Sealed and competitive quotations are invited for the supply of Autoclave (1 No.) for the bio control laboratory of this station. The details of the equipment and accessories are furnished below.

Sl. No.	Items	No.	Specifications
1.	Autoclave	1	<ul style="list-style-type: none">• Size in mm - 550*750• Power consumption - 6 KW• Capacity – 184 Liter• Double walled construction - inner body made up of heavy gauge stainless steel and outer body with SS sheet or MS sheet powder coated - lid& ring made up of stainless steel - The lid fitted with pressure gauge, safety valve, steam exhaust valve & vacuum breaker provided safety valve set to below 15 PSI - A silicon rubber gasket is fitted in the ring to ensure steam tight joint- foot operated lifting mechanism for lid opening - Electrically operated - ISI quality immersion type heaters-Supports: three cast iron sturdy legs duly painted.

The rate and the details of equipment with an amount of Rs. 1000/- should be remitted along with the quotation as E.M.D in Cash or Demand Draft in favour of Professor and Head, Cardamom Research Station, Pampadumpara – 685553 payable at **SBI, Nedumkandam** should reach this office addressed to the Professor and Head, Cardamom Research Station, Pampadumpara on or before 10.10.2018 at 12.30 p.m. in sealed covers. The envelope containing the quotations should bear the superscription “Quotation for the supply of Autoclave for Biocontrol Laboratory”. The quotations will be opened in front of the quotationers or their authorized agents on the same day at 2.30 p.m. More details can be had from the office during working hours.

The decision of the Professor and Head will be final in this regard. All rules and regulations concerned with Govt. quotations would apply for the same too.




Professor and Head