



KRISHI VIGYAN KENDRA
KOTTAYAM
KERALA AGRICULTURAL UNIVERSITY
Kumarakom, Kottayam - 686 566



No.KVK/949/2019

Date: 29.01.2020

CORRIGENDUM

Sub: Corrigendum- Technical specifications for Vertical Autoclave- Revised- reg.

Ref: No.KVK/449/2019 dt.20.01.2020 of KVK, Kottayam

This is to inform that the Technical specifications for Vertical Autoclave vide reference cited above has been revised as follows.

Technical Specifications for Vertical Autoclave – Qty: 1 (one) no.

1.	Electrically operated, standard vertical pattern:	Yes
2.	Boiler:	heavy stainless steel
3.	Pedal lifting arrangement lid SS:	Yes
4.	SS perforated work basket:	Yes
5.	Immersion heater having automatic ejection device:	Yes
6.	Pressure gauge steam release cock and safety valve fitted in the lid:	Yes
7.	Operating pressure:	15 lb. per sq.inch
8.	Plug and cord:	Yes
9.	Rating:	6.0kw
10.	Volume litre:	170-180 litres
11.	Inner and outer fully stainless steel:	Yes
12.	Digital microprocessor based temperature controller cum indicator:	Yes
13.	Water level cut of device:	Yes
14.	Pressure cut off switch:	Yes
15.	Warranty:	1 year
16.	Certificate of compliance	Yes



Programme Co-ordinator